


Human MIP-1 α /CCL3 (Macrophage Inflammatory Protein-1 α) Pre-Coated ELISA Kit



USER MANUAL

abeomics
www.abeomics.com

Human MIP-1 α /CCL3

(Macrophage Inflammatory Protein-1alpha)

Pre-Coated ELISA Kit

Catalog No: 90-2103

1 \times 96 well Format (96 tests)

Detection Range: 23.44 – 1500 pg/ml

Sensitivity: < 14 pg/ml

This immunoassay kit allows for the in vitro quantitative determination of Human MIP-1 α /CCL3 concentrations in serum, plasma and other biological fluids.

This kit is for Research Use Only. Not for use in diagnostic/therapeutics procedures.

ABEOMICS, Inc.

9853 Pacific Heights Blvd, STE D.

San Diego, CA-92121

Email: info@abeomics.com

Website: www.abeomics.com

TABLE OF CONTENTS

I.	Background	3
II.	Overview	3
III.	Advantages	3
IV.	Storage	4
V.	Precautions for Use	4-5
VI.	Standard Curve	5
VII.	Reagent Preparation and Storage	7-9
VIII.	Assay Procedure	9-11
IX.	References	11
X.	Troubleshooting	12

I. BACKGROUND

MIP-1 α is a chemokine of the CC subfamily which is inducible in a number of hematopoietic cells, particularly in those involved in adaptive immune responses (macrophages, dendritic cells, and B and T lymphocytes). MIP-1 α signals through the chemokine receptors CCR1 and CCR5. Previous *in vitro* and *in vivo* studies highlighted MIP-1 α as a key response gene up-regulated in normal and neoplastic B cells in response to BCR signaling, and repressed by Bcl-6. MIP-1 α has also been implicated in the regulation of cancer cell growth, angiogenesis and metastasis of different tumors such as melanoma, colorectal cancer, and renal cell carcinoma. It increases cellular migration and expression of matrix metalloproteinase (MMP)-2 in human chondrosarcoma cells. MIP-1 α induced expression of MMP-2 and migration are inhibited by specific inhibitors, and inactive mutants of AMPK, p38 mitogen activated protein kinase (p38 or p38-MAPK), and nuclear factor κ B (NF- κ B) cascades. On the other hand, MIP-1 α treatment demonstrably activated AMPK, p38, and NF- κ B signaling pathways.

II. OVERVIEW

This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti- Human MIP-1 α /CCL3 antibody was pre-coated into 96-well plates. Biotin conjugated anti-human MIP-1 α /CCL3 detection antibody was used. Standards, test samples and biotin conjugated detection antibody were added to the wells subsequently. Wash buffer was used to wash any non-specific binding. HRP conjugated Streptavidin was used as secondary antibody. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Human MIP-1 α /CCL3 amount of samples captured in the plate. Optical Density (O.D) can be read at absorbance 450nm in a microplate reader. Concentration of Human MIP-1 α /CCL3 can be calculated using the standard curve.

III. ADVANTAGES

Multiple samples can be analyzed in a low volume, high-throughput format.

Full analysis can be complete in 4 hours.

IV. STORAGE

Kit can be stored in 4°C, if you are using within a week.

If you are using within 6 months, lyophilized standard can be stored in -20°C and other components at 4°C.

Kit Components

Item	Specifications	Storage
96 well Strip ELISA Plate	8 X 12 well	4°C/-20°C
Lyophilized Standard	2 vials	4°C/-20°C
Sample and Standard Dilution Buffer	20 ml	4°C
Biotinylated Detection Antibody for MIP-1 α /CCL3	120 μ l	4°C/-20°C
Antibody Dilution Buffer	10 ml	4°C
HRP Conjugated Streptavidin (SABC)	120 μ l	4°C in dark
SABC Dilution Buffer	10 ml	4°C
TMB Substrate	10 ml	4°C in dark
Stop Solution	10 ml	4°C
25X Wash Buffer	30 ml	4°C
Plate Sealer	5 pieces	
Product Manual	1	

Material Required, (Not Supplied)

Microplate Reader
 37°C Incubator
 Plate Reader
 Multi Chanel Pipette and disposable tips
 Eppendorf Tubes
 Deionized Water


V. PRECAUTIONS FOR USE

1. To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.


2. After opening and before using, keep plate dry.
3. Before using the Kit, spin tubes and bring down all components to the bottom of tubes.
4. Storage TMB reagents avoid light.
5. Washing process is very important, not fully wash easily cause a false positive.
6. Duplicate well assay is recommended for both standard and sample testing.
7. Don't let Micro plate dry at the assay, for dry plate will inactivate active components on plate.
8. Don't reuse tips and tubes to avoid cross contamination.
9. Avoid using the reagents from different batches together.

VI. STANDARD CURVE

Human MIP-1 α /CCL3 Standard Curve is shown below.



X	pg/ml	1500	750	375	187.5	93.75	46.9	23.44	0
Y	O.D.450	2.098	1.217	0.666	0.362	0.196	0.114	0.071	0.048



VII. REAGENT PREPARATION AND STORAGE

Included buffers and reagents are optimized for use with this kit. Substitution with other reagents is not recommended and may not give optimal results.

1. **Reconstitute the lyophilized Standard:** Standard should be prepared no more than 2 hours before the experiment. Use one tube for each experiment.
 - a. Quick spin down one vial of lyophilized standard. (**DO NOT dilute standard directly on the plate**). Add 1ml of sample/standard dilution buffer into one of the standard tube. Incubate at room temp. for 10 min. Mix thoroughly by vortex. Stock Standard concentration is 1500 pg/ml.
 - b. Label 6 eppendorf tubes with 750 pg/ml, 375 pg/ml, 187.5 pg/ml, 93.75 pg/ml, 46.9 pg/ml, 23.44 pg/ml respectively. Add 0.3 ml of sample/standard dilution buffer into each tube. Add 0.3 ml of stock standard (1500 pg/ml) into 1st tube and mix thoroughly. Transfer 0.3 ml from 1st tube to 2nd tube and mix thoroughly. Transfer 0.3 ml from 2nd tube to 3rd tube mix thoroughly, and so on.




Fig-1: Dilution tubes

Note: Standard Solutions are best used within 2 hrs. Standard solution should be stored at 4°C for up to 12 hrs. or store at -20°C for up to 48 hrs. Avoid repeated freeze-thaw.

2. **Sample Preparation and storage:** Test samples should be collected, analyze immediately (within 2 hrs.) or aliquot and store at -20°C for long term. Avoid multiple freeze-thaw cycles.
 - a. **Cell culture supernatants:** Centrifuge to remove precipitate, analyze immediately or aliquot and store at -20°C.

- b. Serum:** Coagulate the serum at room temp about 1 hr. Centrifuge approximately 1000 \times g for 15 min. Analyze serum immediately or aliquot and store at -20°C.
- c. Plasma:** Collect plasma with heparin or EDTA as the anti-coagulant. Centrifuge for 15 min at 2-8°C at 1500 \times g within 30 min of collection. For eliminating the platelet effect, suggesting that further centrifugation for 10 min at 2-8°C at 10,000 \times g. Analyze immediately or aliquot and store frozen at -20°C.
- d. Tissue Homogenates:** For general information, hemolytic blood may affect the results, you should rinse the tissues with ice cold PBS (0.01M, pH 7.4) to remove excess blood thoroughly. Tissue pieces should be weighed and then minced to small pieces. This will be homogenized in PBS in a cold glass homogenizer. (*Volume depends on the weight of the tissue, 1gram of tissue requires 9 ml of ice cold PBS with protease inhibitor*). To further break the cells, you can sonicate the suspension with an ultrasonic cell disrupter or subject it to freeze-thaw cycle. Homogenates are then centrifuged for 5 min. at 5000 \times g to get the supernatant.

Note: *Samples to be used within 5 days may be store at 4°C, otherwise sample should be stored at -20°C (< 1 month) or -80°C (< 2 months) to avoid loss of bioactivity and contamination. Hemolyzed samples are not suitable for use in this Assay.*

- e.** End user should estimate the concentration of the target protein in the test samples first, then select proper dilution factor to make the diluted target protein concentration falls the optimal detection range of the kit. Dilute the samples with the provided dilution buffer. Several trials may be necessary in practice. The test sample should be well mixed with the dilution buffer. Standard curve and sample should be made before the experiment.

High target protein concentration 15-150 ng/ml: Dilute 1:100 (add 1 μ l of sample into 99 μ l of sample/ standard dilution buffer).

Medium target protein concentration 1.5-15 ng /ml: Dilute 1:10 (add 10 μ l of sample into 90 μ l of sample / standard dilution buffer).

Low target protein concentration 23.48-1500 pg/ml: Dilute 1:2 (add 50 μ l of sample into 50 μ l of sample / standard dilution buffer).

Very low target protein concentration < 23.44 pg/ml: Do not dilute, use 100 μ l of sample.

- 3. Preparation of Biotin detection antibody working solution:** Prepare within one hour before the experiment. Calculate total volume working solution required. (0.1 ml/well \times number of wells. Add 100-200 μ l extra).

Dilute Biotin detection antibody with antibody dilution buffer at 1:100 and mix thoroughly. (*i.e.* add 1 μ l of Biotin conjugated detection antibody into 99 μ l of antibody dilution buffer).

- 4. Preparation of HRP-Streptavidin Conjugate (SABC) working solution:** Prepare within 30 min before the experiment. Calculate total volume working solution required. (0.1 ml/well \times number of wells. Add 100- 200 μ l extra).

Dilute SABC with SABC dilution buffer at 1:100 and mix thoroughly. (*i.e.* add 1 μ l of SABC into 99 μ l of SABC dilution buffer).

- 5. Preparation of 1 X Wash buffer:** Prepare 1 X Wash buffer by diluting 25X Wash buffer in sterile water. Diluted Wash buffer may be stored at 4 $^{\circ}$ C, however we recommend preparing fresh 1X wash buffer for each experiment.

For example: 10 ml of 25X Wash buffer in 240 ml of sterile water.

VIII. ASSAY PROCEDURE

Before starting the experiment, equilibrate the SABC working solution and TMB substrate for at least 30 min at room temp. When diluting samples and reagents, they should be mixed completely and evenly. It is recommended to plot a standard curve for each test.

** If not all microplate strips will be used, remove the excess strips by pressing up from underneath each strip. Place excess strips back in the foil pouch with the included desiccant pack and reseal.*

1. Set standard, test sample and blank (control zero) wells on the pre-coated plate and then record their position. It is recommended to measure each standard and sample in

- duplicate. **Note:** Wash plate twice before adding standard, sample and blank into the well.
2. Add 0.1 ml of standard (1500 pg/ml, 750 pg/ml, 375 pg/ml, 187.5 pg/ml, 93.75 pg/ml, 46.9 pg/ml, 23.44 pg/ml, control zero dilution buffer) into standard well.
 3. Add 0.1 ml of diluted samples into test sample wells.
 4. Seal plate with a cover and incubate at 37°C for 90 min.
 5. Remove the cover and discard samples and standard solution by tapping plate on an absorbent paper. **Note:** DO NOT let the wells completely dry any time. DO NOT wash plate.
 6. Add 0.1 ml of Biotin-detection antibody working solution into the above wells (Standards, control zero and samples).
 7. Seal plate with cover and incubate at 37°C for 60 min.
 8. Remove the cover, and wash plate 3 times with 1X wash buffer.
 9. Add 0.1 ml of SABC working solution into each well. Cover the plate and incubate at 37°C for 30 min.
 10. Remove the cover and wash plate 5 times with 1X wash buffer. Each time let the wash buffer stay in the well for 1-2 min.
 11. Add 90 μ l of TMB substrate into each well, cover the plate and incubate at 37°C in dark within 15-30 min. (**Note:** This incubation time is for reference use only. The optimal time should be determined by end user). The shades of blue can be seen in the first 3-4 wells, only on most concentrated standards. Other wells show no obvious color.
 12. Add 50 μ l of stop solution into each well and mix thoroughly. Color will change into yellow immediately.
 13. Read O.D. absorbance at 450 nm in a micro-plate reader immediately after adding the stop solution.
 14. Calculation: Relative O.D. 450 = O.D. for each well – O.D. 450 control zero well. The Standard curve can be plotted as the relative O.D. 450 of each standard solution in Y axis vs. the respective concentration of the standard in X axis. Concentration of the samples can be incorporated from the standard curve. If the samples were diluted, multiply the dilution factor to the concentration.

Table-1

	Standard 1	Standard 2	3	4	5	6	7	8	9	10	11	12
A	1500pg/ml	1500pg/ml										
B	750pg/ml	750pg/ml										
C	375pg/ml	375pg/ml										
D	187.5pg/ml	187.5pg/ml										
E	93.75pg/ml	93.75pg/ml										
F	46.9pg/ml	46.9pg/ml										
G	23.44pg/ml	23.44pg/ml										
H	0 pg/ml	0 pg/ml										

IX. REFERENCES

1. CCL3 (MIP-1 α) plasma levels and the risk for disease progression in chronic lymphocytic leukemia.
PMID: 21115978
2. AMP-activated protein kinase activation mediates CCL3-induced cell migration and matrix metalloproteinase-2 expression in human chondrosarcoma.
PMID: 24047437
3. Autocrine CCL3 and CCL4 induced by the oncoprotein LMP1 promote Epstein-Barr virus-triggered B cell proliferation.
PMID: 23760235

X. TROUBLE SHOOTING

Problem	Probable Cause	Suggestion
No signal	Forgot to add all components.	Prepare check list and add the components in the correct order.
Low signal	Not enough lysates per well.	Check the protein concentration. Add more lysates.
High background	Washing is not sufficient.	Wash plates thoroughly after incubation with Streptavidin-HRP secondary



ABEOMICS, Inc.
9853 Pacific Heights Blvd, STE D.
San Diego, CA-92121
Email: info@abeomics.com
Website: www.abeomics.com